



# 校准报告

CALIBRATION REPORT

报告编号: 173200531  
Report No

第 4 页, 共 5 页  
Page 4 of 5 Pages

## 校准结果

Results of Calibration

1 外观及工作正常性检查: 正常。

Check on Appearance and Function: Normal

2 直流电流测量误差 : 见表 1

Indication Error of DC Current Measurement : See Table 1

表1 Table 1 直流电流输出测量

输入电流 Input Curr.	输出电流理论值 Theo. Curr.Value	电流实测值 Meas. Value	误差 Error	最大允许误差 MPE	结论 Conclusion
(A)	(mA)	(mA)	(mA)	(mA)	(Pass/Fail)
40.000	20.0000	19.9992	-0.0008	± 0.0400	P
60.000	30.0000	29.9998	-0.0002	± 0.0600	P
80.000	40.0000	39.9989	-0.0011	± 0.0800	P
100.000	50.0000	49.9985	-0.0015	± 0.1000	P
120.000	60.0000	59.9980	-0.0020	± 0.1200	P
140.000	70.0000	70.0014	0.0014	± 0.1400	P
160.000	80.0000	79.9996	-0.0004	± 0.1600	P
180.000	90.0000	89.9990	-0.0010	± 0.1800	P
200.000	100.0000	99.9997	-0.0003	± 0.2000	P

Notes:

1 本次校准的测量不确定度说明

Measurement Uncertainty in the Calibration

1.1 依据 JJF 1059.1-2012 测量不确定度评定与表示

According to JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement.